VERIFICATION OF CALIBRATION REPORT
of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100390   *Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*   Date:12/13/2011

External Standard Test Values

```
EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.55: 0.079
LOT #: 02511080A3
EXPIRATION: 02/01/2013

BLANK TEST            0.000   10:46
INTERNAL STANDARD   VERIFIED   10:46
EXTERNAL STANDARD     0.081   10:47
BLANK TEST            0.000   10:48
EXTERNAL STANDARD     0.080   10:48
BLANK TEST            0.000   10:49
EXTERNAL STANDARD     0.080   10:49
BLANK TEST            0.000   10:50
EXTERNAL STANDARD     0.080   10:50
BLANK TEST            0.000   10:51
EXTERNAL STANDARD     0.081   10:51
BLANK TEST            0.000   10:52

Average = 0.0804
Std Dev = 0.0005
```

Diagnostic Check

```
VERSIONS
DMT: 1.04
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C          PASSED
Breath Tube    = 48.1°C          PASSED

PUMP INFO
Flow Rate = 5.503 L/M            PASSED

DETECTOR INFO
PUMP ON                          PASSED
PUMP OFF                         PASSED

FILTER INFO
Filter 1                         PASSED
Filter 2                         PASSED
Filter 3                         PASSED

INTERNAL STANDARD                PASSED
```

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 25 day of Jan, 20 12

Nikki Roth, Notary Public
My Commission Expires With Office